

POLICE



DEPARTMENT

No. C1-11337/2017/FSL

Forensic Science Laboratory,

Thiruvananthapuram

dirfsl.pol@kerala.gov.in

04712721533

Dated. 04-08-2018

QUOTATION NOTICE

Sealed quotations are invited for the ""Laminar Flow Chamber/PCR Work Station (1No)" required for the DNA division of Regional Forensic Science Laboratory,Kannur.

The envelopes containing the quotations should bear the superscription "**Purchase of Laminar Flow Chamber/PCR Work Station (1No)**" and should be addressed to the Director, Forensic Science Laboratory, Police Department, Thiruvananthapuram, Kerala. Intending firms may submit their quotations on their own papers. The last date for the receipt of quotations is 18.08.2018, 2 p.m. Late quotations will not be accepted. The quotations will be opened at Forensic Science Laboratory, Thiruvananthapuram on **18.08.2018 at 3 PM.** in the presence of such of the bidders or their authorized representatives. The maximum period required for the delivery of the articles should be mentioned therein the quotation.

Details of the requirements, specification and the conditions governing their supply can be obtained free of cost on request from the Director, Forensic Science Laboratory, Tvpm till **18.08.2018, 2 p.m.**

Latha Devi M A
Joint Director



C1-11337/2017/FSL



a53708

POLICE



DEPARTMENT

No. C1-11337/2017/FSL

Forensic Science Laboratory,
Thiruvananthapuram
dirfsl.pol@kerala.gov.in
04712721533
Dated. 04-08-2018

LAMINAR FLOW CHAMBER (PCR WORK STATION)specification

Inner dimension (WxDxH)	: 120 cm x 55cm x 55 cm or similar range
Side window	: UV absorbing tempered glass or similar
Main Body	: Stainless Steel/ Epoxy coated Steel
Pipette Stand	: Pipette stand inside the chamber desirable
Work surface	: Stainless Steel
Air Filter	: HEPA Filter with efficiency level 99.95% or above
Pre-filter	: washable pre-filter preferred
Stand	: Required Stand should be supplied for the installation of the instrument with proper working height
Main Power Supply	: 230 V 50 Hz
Warranty	: Three years or above

Latha Devi M A
Joint Director



C1-11337/2017/FSL



26ec43