

POLICE



DEPARTMENT

No. C1-12576/2017/FSL

Forensic Science Laboratory,

Thiruvananthapuram

dirfsl.pol@kerala.gov.in

04712721533

Dated. 20-01-2018

CORRIGENDUM

CORRIGENDUM

This is for information of all the bidders that the following amendments are being made in the Tender Notification issued vide Tender No 17/2017 Dtd 8.1.2018 of State Forensic Lab Thiruvananthapuram for the selection of supply of **Digital Zoom Stereo Microscope(2 Nos)**

The bidders are advised to take into account the following amendments in the Specification of the above item before submission of bids against the Tender cited. For convenience and clarity the revised specification is being published in the official Website of Police Department.

Revised Specification for the supply of Digital Zoom

Stereo Microscope

Digital Zoom Stereo Microscope	
Total magnification	Zoom 10X to 70X or above
Working Distance	97 mm or better
Field of view	24 mm or above
Eyepieces	10 X
Objectives	0.6:5 X or better objectives
Viewing Head	Binocular, 35° Inclined Inter pupillary distance: 50 mm – 75 mm
Stage	Clear glass stage plate, 3-3/4" (Φ95 mm) White/Black stage plate, 3-3/4" (Φ95 mm)
Focusing	Adjustment knobs on each side Range: 54 mm – 154 mm




C1-12576/2017/FSL



514572

Illumination	Incident : LED with own switch/intensity knob Transmitted: LED with own switch/intensity knob
Power Supply	100V – 240V, 50/60Hz
Warranty	Three years from the date of Installation
AMC	Should be quoted for the three years after warranty



Dr. R Sreekumar
Director FSL

