

POLICE



DEPARTMENT

No. C1-8821/2020/FSL

Forensic Science Laboratory,

Thiruvananthapuram

✉dirfsl.pol@kerala.gov.in

☎04712721533

Dated. 23-01-2025

Corrigendum for tender

This is for information of all the bidders that the following amendments are being made in the Tender Notification issued vide Tender No 11/2024-2025 Dtd 03.1.2025 for the "Purchase of Benchtop capillary Electrophoresis System for Human identification (DNA sequencer)".The bidders are advised to take into account the following amendments in the Tender Notice of the above items before submission of bids against the Tender cited. For convenience and clarity the revised CORRIGENDUM is being published in the official Website of Police Department.

1.The Opening date of tender extended upto 31/1/2025

2.Firms are also request to forward the Physical copy of the same.

3.In the specification 2,6 and 20 points may be revised as

2.The system (s) should be able to perform sequencing and fragment analysis.

6.Auto sampler capacity: Must have automated sample injection system compatible with two or more 96-well plates/stripes.

20.IQ-OQ certification should be provided during installation and for subsequent periods.

PRADEEP SAJI K
Director FSL



C1-8821/2020/FSL



2f843e