POLICE



DEPARTMENT

No. H9-55509/2020/PHQ(1)

Police Headquarters, Thiruvananthapuram Mphq.pol@kerala.gov.in 04712721547 Dated. 11-08-2020

Quotation Notice No.47/2020

Sealed quotations with samples are invited for the supply of Laminar Airflow Chamber . Intending vendors may quote their offers separately.

| Name of the Item | No. of item |
|-------------------------|-------------|
| Laminar Airflow Chamber | 1 |

Laminar Air Flow Chamber –

Technical specification

| Material of construction : Steel powder coated Work top : Stainless steel good quality Side panels : Transparent acrylic framed in anodised aluminium frame Blowers : Good quality noiseless and vibration less Filtration : Pre Filter <95 micron and HEPA Filter <99.97micron Air velocity : 90 ft / min ±20% UV Tube Manometer Gascock Onsite DOP testing of filtration quality |
|--|
| 3)Side panels : Transparent acrylic framed in anodised aluminium frame 4)Blowers : Good quality noiseless and vibration less 5)Filtration : Pre Filter <95 micron and HEPA Filter <99.97micron 6)Air velocity : 90 ft / min ±20% 7)UV Tube 8)Manometer |
| 4)Blowers : Good quality noiseless and vibration less 5)Filtration : Pre Filter <95 micron and HEPA Filter <99.97micron 6)Air velocity : 90 ft / min ±20% 7)UV Tube 8)Manometer |
| 5)Filtration : Pre Filter <95 micron and HEPA Filter <99.97micron 6)Air velocity : 90 ft / min ±20% 7)UV Tube 8)Manometer |
| 6)Air velocity : 90 ft / min ±20% 7)UV Tube 8)Manometer |
| 7)UV Tube 8)Manometer |
| 8)Manometer |
| |
| 9)Gascock Onsite DOP testing of filtration quality |
| |
| 10)Working chamber size :120 x 60 cm |
| 11)Certifications : |
| 1. CE certified |
| 2. ISO 13485: 2003 (Medical devices) |
| 3. ISO 9001: 2015 |
| 12)Documentation : IQ, OQ & PQ with Manufacturer's Test reports and |
| User's manual |
| 13) Vertical - with CE Mark |
| 14)Warranty period –minimum 3 years from the date of instillation or more |

The quotations should contain price of the item including taxes, guarantee/warranty etc. The quotation complete in all respect should be submitted in envelope bearing the name of the item with samples and should be addressed to State Police Chief ,Kerala,PHQ,TVPM. Intending Tenderers may submit their quotation in their own letter head along with TIN/CST/GSTIN Number. Last date for receipt of quotation is **25.08.2020**





at 2 PM. Late quotations will not be accepted. The quotations will be opened at 3 PM on the same day at PHQ in the presence of the Tenderers or their authorized representatives who may be present at that time. Payment will be made only after the verification of the supplied items. Advance payment will not be allowed.

1. Saxen

VAIBHAV SAXENA IPS Additional Assistant Inspector General For State Police Chief

То

: Director,FSL for necessary action to circulate the notice among known vendors.





