

POLICE



DEPARTMENT

No. C1-5306/2021/FSL

Forensic Science Laboratory,

Thiruvananthapuram

✉dirfsl.pol@kerala.gov.in

☎04712721533

Dated. 06-10-2021

Corrigendum for Laboratory Refrigerator

This is for information of all the bidders that the following amendments are being made in the Tender Notification issued vide Tender No 09/2021-2022 Dtd 05.10.2021 for the "**Purchase of Laboratory Refrigerator**".The bidders are advised to take into account the following amendments in the Tender Notice of the above item before submission of bids against the Tender cited. For convenience and clarity the revised CORRIGENDUM is being published in the official Website of Police Department.

Specification

- 1.Temperature range :1 degree C~ 8 degree C
- 2.Capacity not to be lesser than 1000 Lit
- 3.Minimum 5 Nos. of shelves (Adjustable) required.
- 4.Double door preferred
- 5.Facility for locking the doors
- 6.Microprocessor controller
- 7.Digital temperature display
- 8.Equipment should be frost free/static cooling
- 9.Forced air cooling required
- 10.White exterior preferred
- 11.Equipment should visual & Acoustic alarms
- 12.Alarms should available for high/low temp,door open etc.
- 13.CFC free refrigerator
- 14.PUF insulation
- 15.Lockable Castor wheels
- 16.Suitable voltage stabilizer should provide
- 17.Minimum three year warranty



C1-5306/2021/FSL



b13cf1

Jithediv

Latha Devi M A
Director FSL

Copy To : 1.JD,RFSL,Kannur 2.JD,RFSL,TSR



C1-5306/2021/FSL



b13cf1