

POLICE



DEPARTMENT

No. C1-2892/2026/FSL

Forensic Science Laboratory,
Thiruvananthapuram
PHQ Compound,
Vazhuthacaud
Pincode:695010

✉dirfsl.pol@kerala.gov.in

☎04712721533

Dated. 05-05-2026

Quotation Notice for Centrifuge

Quotation number	06/2026-27
Due date and time for receipt of quotation	13/05/2026 12.NOON
Date, Place and Time of opening of quotations	13/05/2026 3.00PM,Forensic Science Laboratory, Thiruvananthapuram
Designation and address of officer to whom the quotation is to be addressed	Director, Forensic Science Laboratory, Thiruvananthapuram-695010

Sealed quotations are invited for the " Purchase of non refrigerated centrifuge for Biology Division of Forensic Science Laboratory,Thiruvananthapuram". The envelopes containing the quotations should bear the superscription "Purchase of non refrigerated centrifuge for Biology Division of Forensic Science Laboratory,Thiruvananthapuram" and should be addressed to the Director, Forensic Science Laboratory, Police Department, Thiruvananthapuram 695010, Kerala. The last date for the receipt of quotations is 13/05/2026 ,12.NOON. Late quotations will not be accepted. The quotations will be opened at Forensic Science Laboratory, Thiruvananthapuram on 13/05/2026 , 3.00 PM. in the presence of such of the quotationer or their authorized representatives. The maximum period required for the delivery of the articles should be mentioned. Details of the requirements and the conditions governing their supply can be obtained free on request from the Director, Forensic Science Laboratory, Tvpm till 13/05/26 12.NOON.

Specification:

Non refrigerated Centrifuge

- 1.Brushless Induction motor
- 2.Speed control through VFD
- 3.Microprocessor controller



C1-2892/2026/FSL



074c2f

4. Feather touch keys
5. Smooth and soft start
6. Green LED Display
7. Lid interlock
8. Dynamic brake
9. Imbalance detection
10. Emergency lid opening
11. Accurate speed ± 10 RPM
12. Timer up to 99 minutes
13. Last parameter recall
14. Calibration window
15. Elegant lid and front panel
16. Choice of swing out
17. Maximum Speed : 8000 rpm
18. Maximum RCF: 2380
19. Tube capacity: 15 ml
20. No. of wells: 16
21. NABL Calibration certificate



PRADEEP SAJI K
Director FSL

Copy To : AD)B) for information

